


# Analysis of Errors Made During 163 IMRT Irradiations of an Anthropomorphic Phantom


**Andrea Molineu, Paola Alvarez,  
Nadia Hernandez, David S. Followill,  
Geoffrey S. Ibbott**




# RPC Phantoms




**prostate RTOG 0126 (IMRT)**



**thorax RTOG 0236 (SBRT)**




**H&N IMRT  
RTOG 0225, 0126;  
COG ACNS0331**



**liver RTOG 0438**

# IMRT H&N Phantom

- **Primary PTV**  
4 cm diameter  
4 TLD
- **Secondary PTV**  
2 cm diameter  
2 TLD
- **Organ at risk**  
1 cm diameter  
2 TLD
- **Axial and sagittal radiochromic films**




- **1° PTV treated to 6.6 Gy**
- **2° PTV treated to 5.4 Gy**
- **OAR limited to < 4.5 Gy**

Designed in collaboration with RTOG;  
Molineu et al, IJROBP, October 2005

# Criteria for credentialing

- **RPC/Inst dose in PTVs: 0.93-1.07**
- **distance to agreement in high gradient region near OAR:  $\leq 4$  mm**



# Phantom Results

<b>Phantom</b>	<b>H&amp;N</b>	<b>Prostate</b>	<b>Thorax</b>	<b>Liver</b>
<b>Irradiations</b>	<b>163</b>	<b>27</b>	<b>17</b>	<b>-</b>
<b>Pass</b>	<b>115*</b>	<b>24</b>	<b>15</b>	<b>-</b>
<b>Fail</b>	<b>48</b>	<b>3</b>	<b>2</b>	<b>-</b>
<b>Under analysis or at institution</b>	<b>10</b>	<b>3</b>	<b>5</b>	<b>2</b>
<b>Year introduced</b>	<b>2001</b>	<b>Spring 2004</b>	<b>Spring 2004</b>	<b>Spring 2005</b>

# IMRT H&N Phantom Results

- 163 irradiations were analyzed
- 115 irradiations passed the criteria
  - 28 institutions irradiated multiple times
- 48 irradiations did not pass the criteria
- 128 institutions are represented


**Only 68% of institutions passed the criteria on the first irradiation.**

# IMRT H&N Phantom Results cont.

- **28 failed by absolute dose only**
- **7 failed by DTA only**
- **13 failed by both absolute dose and DTA**


	<b>1PTV</b>	<b>2PTV</b>	<b>DTA (mm)</b>
mean	0.99	0.99	-0.7
std dev	0.078	0.065	3.5
count	450	223	162
range	0.78-1.13	0.92-1.22	-15 to +8

# Physicists per machine





# Number of Machines



# Results grouped by accelerator manufacturer

Linear Accelerator Manufacturer	Pass Rate (%)	Attempts	Criteria Failed		
			Dose	DTA	Dose and DTA
BrainLab	100	1	0	0	0
Elekta	54	13	5	1	0
Siemens	66	29	4	1	5
TomoTherapy	67	3	1	0	0
Varian	74	117	18	5	8
total		163	28	7	13

# Results grouped by TPS

Treatment planning system	Pass Rate (%)	Attempts	Criteria Failed		
			Dose	DTA	Dose and DTA
BrainScan	100	4	0	0	0
Cadplan	67	3	1	0	0
CMS XiO	76	17	1	1	2
Corvus	73	26	6	0	1
Eclipse	84	32	2	2	1
Helax	100	2	0	0	0
Pinnacle	61	69	16	4	7
Radionics XKnife	100	1	0	0	0
Theraplan Plus	0	2	0	0	2
TomoTherapy	67	3	1	0	0
Inst. developed TPS	75	4	1	0	0
total		163	28	7	13

# Results grouped by machine/TPS combo

Manufacturer/TPS Combination	Pass Rate (%)	Attempts	Criteria Failed		
			Dose	DTA	Dose and DTA
Elekta/Corvus	0	1	1	0	0
Elekta/Eclipse	--	0	0	0	0
Elekta/Pinnacle	55	11	4	1	0
Elekta/XiO	100	1	0	0	0
Siemens/Corvus	88	8	1	0	0
Siemens/Eclipse	--	0	0	0	0
Siemens/Pinnacle	54	13	2	0	4
Siemens/XiO	50	4	0	1	1
Varian/Corvus	71	17	4	0	1
Varian/Eclipse	84	32	2	2	1
Varian/Pinnacle	64	45	10	3	3
Varian/XiO	83	12	1	0	1
<b>total</b>		144	25	7	11

# Explanations for Failures

Explanation	Minimum # of occurrences
incorrect output factors in TPS	1
incorrect PDD in TPS	1
inadequacies in beam modeling at leaf ends (Cadman, et al; PMB 2002)	14
not adjusting MU to account for dose differences measured with ion chamber	3
errors in couch indexing with Peacock system	2
2 mm tolerance on MLC leaf position	1
setup errors	7
target malfunction	1

# Changes made by institutions that resulted in acceptable phantom irradiation

## Changes

**input new output factors  
remeasured PDD and modeled  
beam based on new values  
adjusted leaf end modeling  
updated software version  
upgraded MLC leaves  
more accurate setup  
replaced target**

# Conclusions

- The RPC phantom provides a comprehensive evaluation of IMRT for clinical trials
- QA of IMRT is important!



The investigation was supported by PHS grants CA10953 and CA81647 awarded by the NCI, DHHS.