

Vendor Data Submission Status

- Accuray CyRIS MultiPlan v1.5.2 – P0236 phantom dataset received from clinical site. Looks good, but deemed inadequate for compliance testing (single fraction treatment)
- CMS XiO 4.3.3 ATC compliant (without footnote)
- Nucletron Oncentra MasterPlan v1.5 ATC compliant
- Philips Pinnacle³ v8.0h is in “limited release” with DICOM export. ITC & Philips looking for test site. Last testing with vendor was good (6/19/2007)

Vendor Data Submission Status

- Prowess Panther 4.40 dataset received from clinical site (complete 7/25/2007). Not yet reviewed.
- Tomotherapy v3.0 (already posted as compliant) is now shipping.
- Varian Eclipse still does NOT export interpolated planes of image data. Therefore structure contours do not align when interpolation is done (QARC Benchmark dataset shows this)