

Vendor Data Submission Status

- Philips Pinnacle³ v8.0h – Posted as ATC compliant 5Mar2008
- Accuray CyRIS MultiPlan v1.5.2 – Posted as ATC compliant 22Aug2007
- Prowess Panther 4.41 – Posted as ATC compliant for 3DCRT, IMRT, & Seed brachy 11Nov2007

Vendor Data Submission Status

- CMS XiO 4.3.3 – HFP dataset with incorrectly oriented dose array received 3/11/2008 (TJU)
- Nucletron Oncentra MasterPlan
 - v3.1 test data received from vendor 3/13/2008 (Tele and HDR)
 - No major problems
- Siemens KonRad – 3/11/2008 Clinical site (St. Jude) data received
 - Dose is coarse in space and value (??)
 - 7/19/2004 Last vendor contact (similar dose problem then)

Vendor Data Submission Status

- **SonoTECH**
 - 3/21/2008 - Email contact, SFTP, and RRT established
 - Small German Company (brachy only?)
 - Customers want to participate in protocols
- **US dataset – 3/26/2008 (William Beaumont)**
 - Originally sought for non-square pixels
 - Has other unexpected “problems”